

CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Answer Precision Technologies, Inc. 150 Bullock Street Kitchener, ON N2C 0C5 Canada

(and the satellite location as listed on the scope)

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

DIMENSIONAL MEASUREMENT

This certificate is valid only when accompanied by a current scope of accreditation document. The current scope of accreditation can be verified at <u>www.anab.org</u>.







Jason Stine, Vice President

Expiry Date: 10 April 2026 Certificate Number: L2384

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Answer Precision Technologies, Inc

150 Bullock Street Kitchener, ON N2C 0C5 Dan Andrews 519-748-0079

DIMENSIONAL MEASUREMENT

Valid to: April 10, 2026

Certificate Number: L2384

3 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ¹	Reference Standard, Method, and/or Equipment
Dimensional Measurement 3D	X = (0 to 1 200) mm Y = (0 to 2 000) mm Z = (0 to 1 000) mm	(6.7 + 38 <i>L</i>) μm	
	X = (0 to 1 200) mm Y = (0 to 2 000) mm Z = (0 to 1 000) mm	(17.4 + 33 <i>L</i>) μm	Coordinate Measuring Machines utilized as Reference Standards for Dimensional Measurements
	X = (0 to 2 500) mm Y = (0 to 5 000) mm Z = (0 to 1 800) mm	(4 + 41 <i>L</i>) μm	





Services performed at satellite location

Answer Precision Mexico S.A. de C.V.

Carretera Estatal 431 KM 2.2, Número exterior: 19, Número interior: módulo 1 Parque Tecnológico Innovación Querétaro El Marqués, Querétaro, Mexico, CP: 76246

3 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ¹	Reference Standard, Method, and/or Equipment
Dimensional Measurement 3D	X = (0 to 1 500) mm $Y = (0 to 2 000) mm$ $Z = (0 to 1 000) mm$	(9.2 + 37 <i>L</i>) μm	Coordinate Measuring Machines utilized as Reference Standards for Dimensional Measurements

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (k=2), corresponding to a confidence level of approximately 95%.

- Notes:
- 1. *L* is length in unit of meter.
- 2. This scope is formatted as part of a single document including Certificate of Accreditation No. L2384.

Jason Stine, Vice President



www.anab.org